



# Certificate of Compliance

**Certificate:** 70009230

**Master Contract:** 162029

**Project:** 70009230

**Date Issued:** September 12, 2014

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

Attention: Thomas Smith

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Cindy Johnson*

**Issued by:** Cindy Johnson, ASCT.

## **PRODUCTS**

**CLASS 8721 06** - LABORATORY EQUIPMENT - Electrical

**CLASS 8721 86** - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Automated Pressure Calibrator, Fluke Calibration Model 4322:

- Pressure Calibrator: Rated 100-240 V, 50-60 Hz, 110 VA.
- Intensifier: Rated 100-240 V, 50-60 Hz, 250 VA.

Notes:

1. The above model is cord connected, Equipment Class I, Pollution Degree 2.
2. Mode of operation: Continuous.
3. Environmental Conditions: Humidity 5 to 95% rH, 0°C to 50°C, 2000 m.



**Certificate:** 70009230

**Master Contract:** 162029

**Project:** 70009230

**Date Issued:** September 12, 2014

---

#### CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.

#### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Reference Standards:

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements