

Manual Supplement

Manual Title:	5520A PQ Option	Supplement Issue:	1
Part Number:	1608366	Issue Date:	7/01
Print Date:	May 2001	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 3, Table 2, 33 V to 329.99 V Composite Waveform Ranges, change the Harmonic Phase Uncertainty (Relative to Fundamental)²

From: 1.0°

To: 1.2°

On page 41, Table 16, under the 60 Hz Frequency for 329.99V Range, change the Specification (deg) for Harmonic 8,

From: 1.0

To: 1.2

On page 42, Table 16, replace the first section of the table with the following:

Verification Tests for AC Voltage		Harmonic	Fundamental	Phase	Amplitude (V)	Specification (V)	Specification (deg)
Range	329.99 V Normal Voltage Output	rms			140.0000	290.00 mV	0.75
Wave	IV	1	100.00%		50.985	122.58 mV	0.75
RMS Output	140 V	2	100.00%	0	50.985	122.58 mV	0.75
Frequency	50 Hz	3	100.00%	0	50.985	122.58 mV	0.75
		4	100.00%	0	50.985	122.58 mV	0.75
		5	100.00%	0	50.985	122.58 mV	0.75
		6	100.00%	0	50.985	122.58 mV	0.75
		7	100.00%	0	50.985	122.58 mV	0.75
		8	30.00%	0	15.296	122.58 mV	0.75
		9	30.00%	0	15.296	153.23 mV	1.2
		10	30.00%	0	15.296	153.23 mV	1.2
		11	30.00%	0	15.296	153.23 mV	1.2
		12	30.00%	0	15.296	153.23 mV	1.2
		13	30.00%	0	15.296	153.23 mV	1.2

On page 43, Table 16, under the 50 Hz Frequency for 329.99 V Range, change the Specification (deg) for Harmonic entries 9, 10, 11 & 12,

From: 1.0

To: 1.2