



## EU DECLARATION OF CONFORMITY

Product Identification:

**Fluke Calibration 5730A - Calibrator**

First declared: 2013

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2014/30/EU, Electromagnetic Compatibility (EMC)**

**Directive 2014/35/EU, Low Voltage (LVD)**

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61326-2-1: 2013 Particular EMC requirements for sensitive test and measurement equipment

EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General

EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits

-----

Special Conditions:

None

Manufacturer

Fluke Calibration  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160603130603 - This declaration supersedes all previous declarations for this product.

03 June 2016