

Manual Supplement

Manual Title:	5790B Operators	Supplement Issue:	1
Print Date:	June 2015	Issue Date:	11/16
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 363

On page 6-53, under Response, remove:

- 4 = Measurement is unsettled ("U" annunciator is lit)

On page 6-71, under VAL?, replace the Response with:

Response

1. (Float) The magnitude of the measurement in volts.
2. (Float) The frequency of the measurement in Hertz.
3. (Integer) A code describing the measurement as follows. If more than one conditions is true, the highest number only is returned.
 - 0 = Measurement conditions valid, no other condition below is true
 - 1 = Frequency underrange
 - 2 = Frequency overrange
 - 3 = Measurement settled, but digital filter not full
 - 4 = Measurement is unsettled
 - 5 = Value is underrange ("Under Range" on display)
 - 6 = Value is overrange ("Over Range" on display)
 - 7 = Value is invalid ("-----" on display)