

# Manual Supplement

Manual Title:	6330 Users Guide	Supplement Issue:	<b>3</b>
Print Date:	January 2013	Issue Date:	12/17
Revision/Date:		Page Count:	1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.  
This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	6330
CD Rev. & Date:	1/2013
CD PN:	3751028

## Change #1

On page 10, after the last bullet in the **Environmental Conditions** section, add:

*Note*

*The Product can show some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m to a maximum of 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.*

## Change #2, 452

On page 9, under **Specifications** change:

From: Digital Setting Accuracy       $\pm 0.5$  °C

To:    Digital Setting Accuracy       $\pm 1.0$  °C

## Change #3, 534

On page 11, under **Specifications**, replace the **Power** row with:

**Power†**    115 V ac ( $\pm 10$  %), 50/60 Hz, 7 A [230 V ac ( $\pm 10$  %), 50/60 Hz, 3.5 A], 900 W