

Manual Supplement

Manual Title:	7320 Users Guide	Supplement Issue:	3
Part Number:	3720888	Issue Date:	12/17
Print Date:	February 2013	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1

On page 12, following the **Environmental Conditions** section, add:

Note

The Product may exhibit some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m up to 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.

Change #2, 517

On page 12, under the **Environmental Conditions**, replace the 1st and 2nd bullets with:

- temperature range: 5 °C to 35 °C (41 °F to 95 °F)
- ambient relative humidity: maximum 80 % for temperatures <31 °C decreasing linearly to 50 % at 35 °C

Change #3, 535

On page 11, under **Specifications**, replace the **Power** row with:

Power 115 V ac (±10 %), 60 Hz, 15 A [230 V ac (±10 %), 50 Hz, 8 A optional], 1400 VA

⚠ Caution

If the voltage is outside ±10 %, the compressor may be damaged. Check the back panel label for the correct voltage and frequency prior to energizing the unit.