

Manual Supplement

Manual Title:	7380 Users Guide	Supplement Issue:	1
Part Number:		Issue Date:	10/17
Print Date:	January 2013	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 529

On page 12, under the **Environmental Conditions**, replace the 1st and 2nd bullets with:

- temperature range: 5 °C to 35 °C (41 °F to 95 °F)
- ambient relative humidity: maximum 80 % for temperatures <31 °C decreasing linearly to 50 % at 35 °C