



EU DECLARATION OF CONFORMITY

Product Identification:

Fluke Calibration 7526A - Precision Process Calibrator

First declared: 2012

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-1: 2013 Particular EMC requirements for sensitive test and measurement equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits

Special Conditions:

None

Manufacturer

Fluke Calibration
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160606123656 - This declaration supersedes all previous declarations for this product.

06 June 2016