

Manual Supplement

Manual Title:	910 & 910R Operators	Supplement Issue:	1
Print Date:	May 2002	Issue Date:	1/18
Part Number:	1580336	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 556

On page 7-2, remove the following Allan deviation Specification:

$$(\tau = 1000 \text{ s}) < 1 \times 10^{-12*} < 5 \times 10^{-11*}$$