



EU DECLARATION OF CONFORMITY

Product Identification:

Fluke Calibration 9114, 9115A, 9116A - Freeze Point Furnace

First declared: 2006

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General

EN 61010-2-010: 2003 Particular requirements for laboratory equipment for the heating of materials

Special Conditions:

None

Manufacturer

Fluke Calibration
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160908074548 - This declaration supersedes all previous declarations for this product.

08 September 2016