

Manual Supplement

Manual Title: 9142, 9143, 9144 Users Guide
Print Date: February 2013
Part Number: 3720986
Revision/Date:

Supplement Issue: **2**
Issue Date: 3/18
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 488

On page 13, in the **Specification** table, replace the **TC Millivolt Range** with:

TC Millivolt Range	-10 mV to 100 mV
---------------------------	------------------

Change #2, 539

On page 12, in the **Base Unit Specifications** table, replace the **Computer Interface** with:

Computer Interface	RS-232
---------------------------	--------