

Manual Supplement

Manual Title: E-DWT-H Operation and Maintenance Supplement Issue: 1
Part Number: Issue Date: 2/15
Print Date: 2009 Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 432

On page 4, replace **Precision**¹ and **Measurement Uncertainty**³ with:

Precision¹

For A200M-B: ± 0.018 % of reading or 0.0036 % of Q-RPT span, whichever is greater.

For all other Q-RPTs: ± 0.018 % of reading or 0.0018 % of Q-RPT span, whichever is greater.

Measurement Uncertainty³

One year for A200M-B: ± 0.02 % of reading or 0.004 % of Q-RPT span, whichever is greater.

One year for all other Q-RPTs: ± 0.02 % of reading or 0.002 % of Q-RPT span, whichever is greater.

Two years for A200M-B: ± 0.025 % of reading or 0.005 % of Q-RPT span, whichever is greater.

Two years for all other Q-RPTs: ± 0.025 % of reading or 0.0025 % of Q-RPT span, whichever is greater.