

Manual Supplement

Manual Title:	PG7000, PG7102, PG7202, PG7302, PG7601	Operation and Maintenance
Print Date:	2011	Supplement Issue: 1
Revision/Date:		Issue Date: 9/15
		Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.



Change #1, 437

On page 4, under the **1.2.1.2 Ambient and Instrument Condition Measurements**, replace the **Vacuum** Resolution with:

Resolution 0.1 Pa