

# Manual Supplement

Manual Title: 2638A Users  
Print Date: June 2013  
Revision/Date: Rev.1, 5/16

Supplement Issue: 1  
Issue Date: 6/21  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 621

On page 1-15, replace the **DC Voltage Additional Errors** table with:

Digits	NPLC	Ch. x01 – x20	Additional NPLC Noise Error
6 ½	200	Add 2 µV	-
6 ½	100	Add 2 µV	-
6 ½	10 (Slow)	Add 2 µV	-
5 ½	1 (Medium)	Add 2 µV	0.0008 % of range
4 ½	0.2 (Fast)	-	0.002 % of range + 12 µV
4 ½	0.02	-	0.014 % of range + 17 µV <sup>[1]</sup>

[1] 0.05% of range for the 300 V range

On page 1-17, replace the **DC Current Additional Errors** table with:

Digits	NPLC	Additional NPLC Noise Error for 10 mA, 100 mA	Additional NPLC Noise Error for 100 µA, 1 mA
6 ½	200	-	-
6 ½	100	-	-
6 ½	10 (Slow)	-	-
5 ½	1 (Medium)	0.0008 % of range	0.0008 % of range
4 ½	0.2 (Fast)	0.002 % of range + 1.2 µA	0.002 % of range + 0.7 µA
4 ½	0.02	0.014 % of range + 1.7 µA	0.014 % of range + 1.0 µA

On page 1-19, replace the **Resistance Additional Errors** table with:

Digits	NPLC	Additional NPLC Noise Error
6 ½	200	0 % of range
6 ½	100	0 % of range
6 ½	10 (Slow)	0 % of range
5 ½	1 (Medium)	0.0008 % of range <sup>[1]</sup>
4 ½	0.2 (Fast) (only for 2-wire)	0.002 % of range + 12 mΩ
4 ½	0.02 (only for 2-wire)	0.01 % of range + 17 mΩ

[1] 0.2 % of measurement for the 10 MΩ range.