

Manual Supplement

Manual Title: 3130 Users
Part Number: 4222409
Print Date: February 2013
Revision/Date:

Supplement Issue: 1
Issue Date: 10/15
Page Count: 1


This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 464

On page 3, add the following to the **Symbols** table:

| Symbol | Description |
|---|--|
|  | Conforms to relevant South Korean EMC Standards. |

On page 45, under **Standards and Agency Approval**, replace Electromagnetic Environment with:

Electromagnetic Compatibility (EMC)

International IEC 61326-1: Portable Electromagnetic Environment

CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

Korea (KCC)..... Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC) 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.