

Manual Supplement

Manual Title:	7007 User Manual	Supplement Issue:	4
Part Number:	3729260	Issue Date:	7/19
Print Date:	October 2013	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 583

On page 7, under **Specifications**, replace the **Temperature Stability** and the **Temperature Uniformity** with:

Temperature Stability	±0.001 °C at 0 °C (water/ethanol) ±0.001 °C at 25 °C (water) ±0.003 °C at 100 °C (oil 5012)
Temperature Uniformity	±0.004 °C at 0 °C (ethanol) ±0.007 °C at 100 °C (oil 5012)

Change #2

On page 7, after the last bullet in the **Environmental Conditions** section, add:

Note

The Product can show some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m to a maximum of 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.

Change #3, 527

On page 7, under the **Environmental Conditions**, replace the 1st and 2nd bullets with:

- temperature range: 5 °C to 35 °C (41 °F to 95 °F)
- ambient relative humidity: maximum 80 % for temperatures <31 °C decreasing linearly to 50 % at 35 °C