

# Manual Supplement

Manual Title:	9141 Users Guide	Supplement Issue:	<b>2</b>
Part Number:	3729494	Issue Date:	10/15
Print Date:	January 2013	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

**FLUKE**®

**Calibration**

## Change #1, 320

On page 4, following the **Cautions**, add:

*Note*

*Depending upon the probe specifications, the use of the heat shield may not be required to protect the hub. Using the heat shield to protect the probe hub may metrologically affect the product performance. A simple piece of aluminum foil also protects the probe hub and does not restrict the air flow as much as the heat shield lessening the metrological impact.*

On page 13, following the **Set-Up**, add:

*Note*

*Depending upon the probe specifications, the use of the heat shield may not be required to protect the hub. Using the heat shield to protect the probe hub may metrologically affect the product performance. A simple piece of aluminum foil also protects the probe hub and does not restrict the air flow as much as the heat shield lessening the metrological impact.*

## Change #2, 406

On page 43, under **10.2 Calibration Procedure**, replace step 1 with:

1. Choose three set-points to use in the calibration of the R<sub>0</sub>, ALPHA, and DELTA parameters. These set-points are generally 100 °C, 350 °C, and 550 °C but other set-points may be used if necessary.