

Manual Supplement

Manual Title: 9142-B-P-156/AF Users Guide
Print Date: May 2013
Part Number: Web-Only
Revision/Date:

Supplement Issue: 1
Issue Date: 9/16
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 488

On page 13, in the *Specification* table, replace the **TC Millivolt Range** with:

TC Millivolt Range	-10 mV to 100 mV
---------------------------	------------------