

PG7000™/RPM™ to PPC™/3990 Interconnections Kit

P/N 3069508

Installation Sheet

Interconnections Kit, P/N 3069508, Includes

DESCRIPTION	QTY	PART NO.
<i>Adaptor, 1/8 in. Tube x 1/8 in. NPTM</i>	<i>2</i>	<i>3068564</i>
<i>Adaptor, 1/8 in. Tube x 1/4 in. NPTM</i>	<i>2</i>	<i>3068641</i>
<i>Adaptor, 1/4 in. Tube x 1/8 in. Tube stub</i>	<i>1</i>	<i>3135767</i>
<i>Plug, 1/8 in. Tube</i>	<i>1</i>	<i>3134224</i>
<i>Tee, 2T</i>	<i>1</i>	<i>3133091</i>
<i>Filter, 2T</i>	<i>1</i>	<i>3134584</i>
<i>Union, 2P</i>	<i>1</i>	<i>3135052</i>
<i>Reducer, 4P x 2P</i>	<i>1</i>	<i>3135076</i>
<i>Tubing, stainless steel, 1/8 in. OD x 100 cm LG</i>	<i>2</i>	<i>3232450</i>
<i>Quick Connect, AN4M</i>	<i>1</i>	<i>3068690</i>
<i>Quick Connect, 1/8 in. NPTF</i>	<i>1</i>	<i>3068708</i>
<i>Quick Connect and Stand Assembly</i>	<i>1</i>	<i>3124512</i>

The 3069508 Interconnection Kit can be used for multiple applications. This installation sheet includes instructions for the following applications:

- ❶ *PG7000/RPM4 as reference connected to PPC for pressure control and quick connect test stand for DUT with inline filter*
- ❷ *PG7000/RPM4 as reference connected to 3990 for pressure control and quick connect test stand for DUT without inline filter*
- ❸ *PG7000/RPM4 as reference connected to 3990 for pressure control and quick connect test stand for DUT with inline filter*

Parts on Test Port or in Accessories of PG7000

<i>Collar</i>	<i>1</i>	<i>3068599</i>
<i>Gland</i>	<i>1</i>	<i>3068428</i>
<i>1/8 in. Tube x DH 200 M Adaptor</i>	<i>1</i>	<i>3069062</i>




Figure 1: To Pneumatically Connect the PG7000 or RPM to the PPC or MPC

Note

If you are using an RPM (not a PG7000) go directly to Step 3b.

- ❶ Remove gland, collar and plastic plug from the PG7000 TEST port.
- ❷ Install gland and collar onto 1/8 in. tube x DH200 male adaptor supplied in PG7000 accessories.

Note

The collar has left-hand threads. Ensure the collar is securely attached. The plastic plug is not used.

- ❸a PG7000 - Screw the adaptor (as assembled) into the PG7000 test port. Torque to 15 Nm (Newton meter).
- ❸b RPM - Screw 1/8 in. tube x 1/8 in. NPT male adaptor into the RPM test port using teflon tape as seal.

- ❹ Connect tee and filter assembly to quick connect stand assembly.
- ❺ Screw 1/8 in. tube x 1/8 in. NPT male adaptor into PPC test port using Teflon tape as a seal.
- ❻ Mount quick connect stand on bench (mounting holes provided).
- ❼ To connect PG7000 or RPM test port to the quick connect stand cut 1/8" stainless steel tubing to desired length and Swage nut together.
- ❽ To connect PPC to the quick connect stand cut 1/8 in. stainless steel tubing to desired length and Swage nut together.

Note

1/8 in. stainless steel tubing can be bent by hand to desired shape but use caution not to kink the tube.




Figure 2: To Pneumatically Connect the PG7000 or RPM to the 3990 without a filter

Note

If you are using an RPM (not a PG7000) go directly to Step ③b.

- ① Remove gland, collar and plastic plug from the PG7000 TEST port.
- ② Install gland and collar onto 1/8 in. tube x DH200 male adaptor supplied in PG7000 accessories.

Note

The collar has left-hand threads. Ensure the collar is securely attached. The plastic plug is not used.

- ③a PG7000 - Screw the adaptor (as assembled) into the PG7000 test port. Torque to 15 Nm (Newton meter).
- ③b RPM - Screw 1/8 in. tube x 1/8 in. NPT male adaptor into the RPM test port using teflon tape as seal.

- ④ Connect tee to quick connect stand assembly using Reducer provided (3135076). Plug one end of the tee using 3134224 plug provided. Item 3135767 can be used in place of the tee if the stand is being placed at the end of the line.
- ⑤ Screw 1/8 in. tube x 1/4 in. NPT male adaptor into 3990 "TO DEVICE UNDER TEST, TEST PORT" using Teflon tape as a seal.
- ⑥ Screw 1/8 in. tube x 1/4 in. NPT male adaptor into 3990 "TO DEADWEIGHT GAUGE, TEST PORT" using Teflon tape as a seal.
- ⑦ To connect PG7000 or RPM test port to the 3990, cut 1/8 in. Stainless Steel tubing to desired length and Swage nut together.
- ⑧ To connect quick connect stand to the 3990, cut 1/8 in. Stainless Steel tubing to desired length and Swage nut together.
- ⑨ Mount quick connect stand on bench (mounting holes provided).

Note

1/8 in. stainless steel tubing can be bent by hand to desired shape but use caution not to kink the tube.




Figure 3: To Pneumatically Connect the PG7000 or RPM to the 3990 with a filter

Note

If you are using an RPM (not a PG7000) go directly to Step 3b.

- ❶ Remove gland, collar and plastic plug from the PG7000 TEST port.
- ❷ Install gland and collar onto 1/8 in. tube x DH200 male adaptor supplied in PG7000 accessories.

Note

The collar has left-hand threads. Ensure the collar is securely attached. The plastic plug is not used.

- ❸ a PG7000 - Screw the adaptor (as assembled) into the PG7000 test port. Torque to 15 Nm (Newton meter).
- ❸ b RPM - Screw 1/8 in. tube x 1/8 in. NPT male adaptor into the RPM test port using teflon tape as seal.
- ❹ Connect tee and filter assembly to quick connect stand assembly.

- ❺ Screw 1/8 in. tube x 1/4 in. NPT male adaptor into 3990 "TO DEVICE UNDER TEST, TEST PORT" using Teflon tape as a seal. Install the 1/8 in. tube plug into the adaptor.
- ❻ Screw 1/8 in. tube x 1/4 in. NPT male adaptor into 3990 "TO DEADWEIGHT GAUGE, TEST PORT" using Teflon tape as a seal.
- ❼ To connect PG7000 or RPM test port to the quick connect stand, cut 1/8 in. Stainless Steel tubing to desired length and Swage nut together.
- ❽ To connect quick connect stand to the 3990, cut 1/8 in. Stainless Steel tubing to desired length and Swage nut together.
- ❾ Mount quick connect stand on bench (mounting holes provided).

Note

1/8 in. stainless steel tubing can be bent by hand to desired shape but use caution not to kink the tube.